

ANAMET REPORT

ANAMET Report 003
November 1996

ANAMET-951
Type-N Measurement Comparison
at Lower RF
-> Some additional Results
using different ANAs

J Ruefenacht

ANAMET reports are produced by, and for, the members of ANAMET. They are intended for fast dissemination of technical information for discussion purposes and do not necessarily represent an official viewpoint. No responsibility is accepted by the author(s) or ANAMET for any use made of the information contained in this report.

Comments on this report should be sent to:

ANAMET
PO Box 92
Malvern
United Kingdom
WR14 3YS

Extracts from this report may be reproduced
provided that the source is acknowledged.

This report has been approved by the ANAMET Steering Committee.

ANAMET-951 TYPE-N MEASUREMENT COMPARISON AT LOWER RF -> SOME ADDITIONAL RESULTS USING DIFFERENT ANAs.

Author: Juerg Ruefenacht, Swiss Telecom PTT

Introduction

During 1995, the Swiss Telecom PTT decided to carry out some additional investigative measurements during participation in the ANAMET-951 exercise. The investigation would enable system-to-system variations in the Swiss Telecom PTT systems to be examined and compared with the ANAMET consensus value produced by the ANAMET comparison exercise.

The ANAMET-951 comparison exercise involved measuring the reflection coefficient of six line terminations fitted with Type N, 50 Ω , coaxial connectors. There were three male and three female line terminations with nominal VSWRs of 1.05, 1.20 and 1.50 to be measured. The participants were invited to submit measurements of the reflection coefficient at the following frequencies; 100 kHz, 300 kHz, 1 MHz, 3 MHz, 10 MHz, 100 MHz and in steps of 100 MHz to 1 GHz.

Measuring systems

The following network analyser systems and setups were used to make the Swiss Telecom PTT internal comparison measurements:

A HP8753D with option 1D5 (High stability frequency reference) with an integrated S-parameter test set. The calibration kit was a HP85032B with selected loads, which have a good return loss up to 6GHz. The "S11 one port" calibration was performed using the internal Open-Short-Load calibration routine over the frequency range 30kHz to 1000MHz.

Main analyser settings:

Test Port power	: 0dBm
IF BW	: 10Hz
SWEEP TIME	: 20s (by measuring a frequency list with 125 points)
AVERAGING	: 16

-> These measurement results were published in the ANAMET-951 final results report.

A HP8510C using the HP8515A S-parameter test set and a HP85054B calibration kit. The "S11 one port" calibration was performed using the internal Open-Short-Load calibration routine over the frequency range 50MHz to 1000MHz.

Main analyser settings:

Power Source 1	: 10dBm
SWEEP TIME	: Not selectable (set by measuring a frequency list with 95 points)
AVERAGING	: 512

A HP3577A using the HP35677A S-parameter test set and a HP35678A calibration kit. The "S11 one port" calibration was performed using the internal Open-Short-Load calibration routine. The measurements were divided in two frequency ranges: 100kHz to 10.1MHz and 10MHz to 200MHz.

Main analyser settings:

AMPTD	: 15dBm
RES BW	: 1Hz
SWEEP TIME	: 30s (by measuring 51 frequency points)
AVG	: NO

A HP8751A network analyser with option 001 (High stability frequency reference) using a HP87511A S-parameter test set and a HP85032 calibration kit with selected loads. The "S11 one port" calibration was performed using the internal Open-Short-Load calibration routine over the frequency range 30kHz to 500MHz.

Main analyser settings:

Test Port power	: 0dBm
IF BW	: 2Hz
SWEEP TIME	: 55s (by measuring a frequency list with 75 points)
AVERAGING	: NO

A HP8751A network analyser with option 001 (High stability frequency reference) using a HP87512A Transmission / Reflection test set and a HP85032 calibration kit with selected loads. The "S11 one port" calibration was performed using the internal Open-Short-Load calibration routine over the frequency range 30kHz to 500MHz.

Main analyser settings:

Test Port power	: 0dBm
IF BW	: 2Hz
SWEEP TIME	: 55s (by measuring a frequency list with 75 points)
AVERAGING	: NO

Comparison results

The next six pages show the results of the internal comparisons. Each measured line termination is presented with a magnitude and a phase difference graph. The phase differences are calculated with the ANAMET Median values as reference.


Conclusions

The results show that the four network analyser types (8753D, 8510C, 3577A and 8751A) have a very good agreement between the different measurements.


All analysers are very sensitive to the selected sweep time (with exception of the 8510C, where the sweep time is not selectable). When making precise measurements it is advisable to reduce the IF bandwidth and to set the sweep time at least to 1s .

Another consideration is the quality of the calibration standards. With the help of these additional measurements, significant differences were found to be dependent on the quality of the fixed loads. This consequence will be discussed in a future ANAMET publication.


Reflection coefficient magnitude
 Type-N (female) Mismatch SN:6753
 nominal VSWR value = 1.05




Reflection coefficient phase-difference from ANAMET median
 Type-N (female) Mismatch SN:6753
 nominal VSWR value = 1.05




Reflection coefficient magnitude
Type-N (female) Mismatch SN:6585
nominal VSWR value = 1.20




Reflection coefficient phase-difference from ANAMET median
Type-N (female) Mismatch SN:6585
nominal VSWR value = 1.20




Reflection coefficient magnitude
Type-N (female) Mismatch SN:6589
nominal VSWR value = 1.50




Reflection coefficient phase-difference from ANAMET median
Type-N (female) Mismatch SN:6589
nominal VSWR value = 1.50




Reflection coefficient magnitude
Type-N (male) Mismatch SN:6791
nominal VSWR value = 1.05




Reflection coefficient phase-difference from ANAMET median
Type-N (male) Mismatch SN:6791
nominal VSWR value = 1.05




Reflection coefficient magnitude
 Type-N (male) Mismatch SN:6808
 nominal VSWR value = 1.20



Reflection coefficient phase-difference from ANAMET median
 Type-N (male) Mismatch SN:6808
 nominal VSWR value = 1.20



Reflection coefficient magnitude
 Type-N (male) Mismatch SN:6467
 nominal VSWR value = 1.50



Reflection coefficient phase-difference from ANAMET median
 Type-N (male) Mismatch SN:6467
 nominal VSWR value = 1.50

