

NATIONAL PHYSICAL LABORATORY

METROLOGY CENTRE

Ref: MOY/SCMI/28 SPECIFICATION OF ACCURACY
(Issue 5)

for

AN AUTOCOLLIMATOR

Type: A TA1 "Microptic" Autocollimator. The autocollimator has a measuring range of 10 minutes of arc and the micrometer drum of the instrument is so graduated that angular deflections of a reflecting surface may be read to 0.1 second of arc. The autocollimator may be rotated through 90 degrees permitting angular measurement in two planes.

Made by: Rank Precision Industries Ltd
Metrology Division

Note: The nomenclature used in this Specification in relation to the field-of-view of the instrument is illustrated in Fig. 1 attached.

LIMITING VALUE OR
MAXIMUM
PERMISSIBLE ERROR

1. **GENERAL**

- 1.1 The workmanship and finish shall be in keeping with a precision instrument of this class.
- 1.2 The instrument shall be marked with an identification number and with the maker's name or trade mark.

2. **BASE**

- | | | |
|-----|--|---|
| 2.1 | The top and side finished faces shall be flat | 0.01 mm |
| 2.2 | The bottom finished face shall be flat. When placed on a truly flat surface the base shall be free from rock | 0.01 mm |
| 2.3 | The top and bottom finished faces shall be parallel | 0.025 mm over the length of the top face. |
| 2.4 | The side finished face shall be square to the bottom face | 0.025 mm over the depth of the side face. |
| 2.5 | The three levelling screws shall operate smoothly. | |
| 2.6 | When the base is directly in contact with a flat horizontal surface the bubble of the circular level shall be approximately central. | |

3. **EYEPIECE**

- 3.1 The focusing movement of the eyepiece shall be smooth.

LIMITING VALUE OR
MAXIMUM
PERMISSIBLE ERROR

3.2 With the eyepiece approximately at the centre of its focusing range it shall be possible to see in focus the microscope graticule setting-lines, the target cross-wires, and the image of the target cross-wires reflected from a truly flat surface *, by viewing them through a low magnification telescope focused for infinity.

* A 50 mm diameter reflector flat to within 0.0001 mm is suitable.

3.3 The range of focusing movement of the eyepiece shall be ±5 dioptries minimum.

4. **MICROMETER**

4.1 The aperture in the cover over the micrometer drum shall be sufficiently large to expose two numbered graduation marks for all positions of the drum.

4.2 The graduation marks on the drum shall be clearly defined.

4.3 The fiducial mark shall be closely the same thickness as the graduation marks.

4.4 The angular value of the scale shall be clearly marked in terms of the corresponding angle of deflection of an observed reflecting surface.

4.5 The feel of the micrometer shall be smooth and even throughout its travel.

4.6 The micrometer shall have an over-run of at least one revolution at each end of its 10-minute range.

5. **MICROSCOPE GRATICULE SETTING-LINES**

5.1 The setting-lines shall be clearly defined and of uniform thickness.

5.2 The spacing of the setting-lines shall exceed the width of the image of the target cross-wires as measured by the micrometer 5 seconds of arc approximately

5.3 The setting-lines shall be square to the microscope transverse line 5 minutes of arc

6. **EXTERIOR MINUTE SCALE**

6.1 The graduation marks shall be clearly defined.

6.2 The graduations of the minute scale shall match the pitch of the micrometer screw 1%

7. **TARGET CROSS-WIRES**

7.1 The target cross-wires shall be clearly defined, straight and of uniform thickness as observed by a visual test.

7.2 Settings made on the extremities of the appropriate target wire image by means of the microscope graticule setting-lines shall agree 1 second of arc as measured on the micrometer drum.

8. **ILLUMINATION**

8.1 The lighting unit shall provide adequate and uniform illumination over the field of view for short working distances and over the illuminated portion for longer working distances. It shall be possible to make satisfactory settings on the target image over distances up to 10 m.

Note: At the larger distances the field of view will naturally be restricted.

9. **TELESCOPE OBJECTIVE**

9.1 The telescope objective lens shall be set accurately for infinity. With the instrument sighted on a flat reflector * placed within 300 mm of the objective lens, a 12 mm aperture positioned centrally with respect to the objective lens aperture shall be traversed ± 6 mm parallel to the micrometer axis without causing the reflected image of the target wire to shift from its original setting by more than

2 seconds of arc.

10. **ACCURACY OF PERFORMANCE**


10.1 When the autocollimator is sighted on a flat reflector *, which has been set square to the top and side finished faces of the base of the instrument, the reflected image of the target cross-wires shall be visible in the field of view.

10.2 With the instrument sighted on a flat reflecting surface * placed within 300 mm of the objective lens, the scale shall be calibrated for progressive and periodic error. The errors of the scale shall be determined at 1-minute intervals throughout its measuring range and at 5-second intervals over each of two revolutions of the micrometer in a central part of the range, and shall meet the following requirements: -

- (i) The standard deviation of the observations during the calibration shall not exceed
- (ii) Calibration errors shall be of a reasonably uniform nature and shall not exceed

0.15 second of arc

2 seconds of arc overall or
0.5 second of arc over any
1-minute interval.

 (Signed)

for Director

