

**EUROMET PROJECT  
FINAL REPORT**

<b>1. Ref. No.:</b> 389	<b>2. Subject Field:</b> Mass	
<b>3. Type of collaboration:</b> Comparison of measurement standards		
<b>4A. Partners:</b> Phase I: CEM, CMA, SP, FI, NPL Phase II: BNM/LNE, CMI, CNR-IMGC, IGM, IPQ, NMi, NPL, OFMET, PTB, SMU, UME		<b>4B. CEC funded?</b> No
<b>5. Participating countries:</b> BE, CH, CZ, DE, DK, ES, FI, FR, GB, IT, NL, PT, SE, SK, TU		
<b>6. Title:</b> 100 MPa hydraulic pressure comparison		
<b>7. Progress:</b>  The measurements for this comparison have been completed and a report agreed and published as "Report on a comparison of pressure standards in the range 10 MPa to 100 MPa", NPL Report CMAM 54.  The results of this comparison were very satisfactory. The conclusion of the report states that all participants who submitted results can be considered as being consistent with the calculated reference value, and with each other, within their claimed uncertainties at a confidence level of k=2.		
<b>8. Coordinator's name:</b> Michael Perkin <b>Address:</b> NPL, Teddington, UK, TW11 0LW <b>Tel:</b> +44 20 8943 6373 <b>Fax:</b> +44 20 8943 6132 <b>Email:</b> <a href="mailto:michael.perkin@npl.co.uk">michael.perkin@npl.co.uk</a>		
<b>9. Completion date:</b> completed	<b>10. Coordinator's signature:</b>	<b>11. Date:</b> 29 September 2000

*Notes for completion of the form overleaf*

\*) Delete as appropriate