NATIONAL PHYSICAL LABORATORY
Time and Frequency Services
Time, Quantum & Electromagnetics Division
Teddington, Middx, United Kingdom TW11 0LW

Web site: www.npl.co.uk/time

_____

N P L   G P S   B u l l e t i n

_____

No.2011-11   November 2011

MJD	Date	[UTC(NPL) - GPS_time] mod 1s (ns)
55866	2011-11-01	0.3
55867	2011-11-02	0.5
55868	2011-11-03	-0.2
55869	2011-11-04	0.2
55870	2011-11-05	1.0
55871	2011-11-06	1.6
55872	2011-11-07	2.2
55873	2011-11-08	2.7
55874	2011-11-09	2.5
55875	2011-11-10	1.6
55876	2011-11-11	0.8
55877	2011-11-12	0.3
55878	2011-11-13	-0.2
55879	2011-11-14	-0.3
55880	2011-11-15	-0.9
55881	2011-11-16	0.1
55882	2011-11-17	-0.3
55883	2011-11-18	-0.5
55884	2011-11-19	-1.3
55885	2011-11-20	-2.2
55886	2011-11-21	-1.7
55887	2011-11-22	-1.6
55888	2011-11-23	-2.0
55889	2011-11-24	-0.2
55890	2011-11-25	-0.9
55891	2011-11-26	-0.4
55892	2011-11-27	-0.3
55893	2011-11-28	-0.1
55894	2011-11-29	-0.3
55895	2011-11-30	-0.7

NOTES:
1. #.# indicates that NPL data are not available.
2. The total 95% confidence interval on each daily value is +/- 22ns.
3. Due to leap seconds, [UTC(NPL) - GPS_time] div 1s = -14ns.
4. UTC(NPL)-GPS_time = [UTC(NPL)-GPS_time] div 1s + [UTC(NPL)-GPS_time] mod 1s.
5. Expressed in words, total difference = leap seconds + column data.
6. This report has been compiled by GPSMONITOR201.EXE version 2.01.
7. The measurements in this report were taken by Dicom GTR50 GPS timing
   receiver s/no 0807183.
8. The measurements in this report are single-frequency C/A code observations
   with the ionospheric delay corrected using a P3 combination of the P1 and P2
   code measurements.
9. No anomalous GPS measurements were detected during the period covered by this
   report.